

## Specification

Diluent for the homogenization of samples for the microbiological examination according to the European Pharmacopoeia Harmonised Method and ISO standard.

## Presentation

10 Prepared bottles  
Bottle 125 ml  
with: 90 ± 3 ml

### Packaging Details

1 box with 10 bottles 125 ml. Injectable cap: Plastic screw inner cap. The use of syringes needles with a diameter greater than 0.8 mm is not recommended.

### Shelf Life

16 months

### Storage

2-25 °C

## Composition

Composition ( g/l ) :

Casein peptone.....	1.0
Sodium chloride.....	4.3
Disodium hydrogen phosphate dihydrate.....	7.2 <sup>(1)</sup>
Potassium dihydrogen phosphate.....	3.6 <sup>(2)</sup>
Histidin.....	1.0
Lecithin.....	3.0
Tween 80 ® .....	30.0

(1) Equiv. to Disodium phosphate dihydrate

(2) Equiv. to Monopotassium phosphate

## Description /Technique

### Description:

This solution is recommended by the European Pharmacopoeia to dilute samples for microbiological examination. The quantity of emulsifying agent used will depend on the amount of fat in the sample being examined.

The addition of the neutralizing agents TLH (Polysorbate 80 - Lecithin - Histidine) may inactivate a variety of disinfectants.

\* The combination of lecithin, polysorbate 80 and histidine neutralizes aldehydes and phenolic compounds.

\* The combination of lecithin and polysorbate 80 neutralizes the quaternary ammonium compounds.

\* The polysorbate 80 neutralizes quaternary ammonium, Iodine, parabens.

\* Lecithin neutralizes Quaternary Ammonium Compounds (QACs), parahydroxybenzoates (parabens), bis-biguanides.

\* Histidine neutralizes formaldehyde.

### Technique:

Use the medium according to intended purposes, samples and validated methods.

**Quality control****Physical/Chemical control**

Color : Yellowish

pH: 7.0 ± 0.2 at 25°C

**Microbiological control**

Prepare tubes / Inoculate ≤10<sup>3</sup> CFU/ tube (productivity)/ Subculture on agar plates, after holding at 20-25°C for 45 min. to 1 h.

Growth Promotion Test 50-100 CFU according to harmonized Pharmacopoeia monographs (EP) and test methods & ISO 11133:2014/A1:2018

Analytical methodology according to ISO 11133:2014/A1:2018; A2:2020.

Aerobic. Incubation at 30-35 °C for 18-72h (bacteria) and 20-25 °C for 3-5 days (moulds and yeast).

**Microorganism**

*Bacillus subtilis* ATCC® 6633, WDCM 00003

*Candida albicans* ATCC® 10231, WDCM 00054

*Escherichia coli* ATCC® 8739, WDCM 00012

*Staphylococcus aureus* ATCC® 6538, WDCM 00032

*Salmonella typhimurium* ATCC® 14028, WDCM 00031

*Ps. aeruginosa* ATCC® 9027, WDCM 00026

*Aspergillus brasiliensis* ATCC® 16404, WDCM 00053

**Growth**

Good. Recovery ±30% T0 (original enumeration)

Good. Recovery ±30% T0 (original enumeration)

Good. Recovery ±30% T0 (original enumeration)

Good. Recovery ±30% T0 (original enumeration)

Good. Recovery ±30% T0 (original enumeration)

Good. Recovery ±30% T0 (original enumeration)

Good. Recovery ±30% T0 (original enumeration)

**Sterility control**

Incubation 48 h at 30-35 °C and 48 h at 20-25 °C: NO GROWTH.

Check at 7 days after incubation in same conditions.

**Bibliography**

- EUROPEAN PHARMACOPOEIA 8.0 (2014) 8th ed. § 2.6.13. Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. EDQM. Council of Europe. Strasbourg.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 16212 Standard (2017) Cosmetics - Microbiology - Enumeration of yeast and mould.
- ISO 21149 Standard (2017) Cosmetics - Microbiology - Enumeration and detection of aerobic mesophilic bacteria.
- ISO 21150 Standard (2015) Cosmetics - Microbiology - Detection of *Escherichia coli*.
- ISO 22717 Standard (2015) Cosmetics - Microbiology - Detection of *Pseudomonas aeruginosa*.
- ISO 22718 Standard (2015) . Cosmetics - Microbiology - Detection of *Staphylococcus aureus*.