

Specification

Liquid medium for fungal isolation and sterility test according to the Pharmacopeial Harmonised Method.

Presentation

10 Prepared bottle
Bottle 125 ml
with: 100 ± 3 ml

Packaging Details

1 box with 10 bottles 125 ml. Injectable cap: Plastic screw inner cap. The use of syringes needles with a diameter greater than 0.8 mm is not recommended.

Shelf Life

16 months

Storage

8-25 °C

Composition

Composition (g/l):

Pancreatic digest of casein.....5.00
Peptic digest of meat.....5.00
Dextrose.....20.0

Description /Technique

Description

This medium is especially adapted to the culture of fungi and acidophilic bacteria.

Sabouraud USP Broth is formulated according to the US Pharmacopoeia, US NF and the 21 CFR guidelines. In the latest editions of these methods Tryptone Soy Broth may be used for sterility checking in pharmaceutical products for injection. This formulation is similar to Antibiotic Medium No. 13 by Grove and Randall and the 21 CFR guideline.

This medium is not selective, but the low pH inhibits the growth of non acidophilic microorganisms. Special measures must be taken while reconstituting and heating the medium due to this strong acid reaction and the high content of glucose. It is important to preheat the autoclave and thereby reach the sterilization temperature as soon as possible otherwise the glucose becomes caramelized turning the medium dark and reducing its effectiveness.

Technique

This medium is recommended for use in many tests and assays, but for a long time has been the medium of choice for verifying the sterility of sterile pharmaceutical products.

The efficacy of the medium and absence of fungistatic products is verified by using *Candida albicans* positive control. A loop of a 1:1000 dilution of a fresh 24 hours culture is added to the control tube and incubated appropriately. Sterility testing must be carried out in a controlled and verified medium. To check fungistatic activity of a product, prepare an control culture inoculum as mentioned above and inoculate two series of tubes of Sabaroud Broth as follows:

- To one set of tubes add a specified amount of product to be tested. This is the test series.
- To another series add only the inoculum. And incubate both series of tubes simultaneously.
- Incubation of both series of tubes must be carried out at 20-22°C for 10 days. After this period compare the results.

If the assay series has less growth than the control series, the product has fungistatic activity. If the growth is equal or more, then it has no fungistatic properties. For the quantitative assay of fungistatic activity, perform the assay with several series of different concentrations until a point of equal growth in both control and test series.

Quality control

Physical/Chemical control

Color : Straw-coloured yellow pH: 5.6 ± 0.2 at 25°C

Microbiological control

Inoculate with 10-100 CFU according to harmonized Pharmacopoeiae or with 10⁴-10⁶ CFU for selectivity.

Analytical methodology according to ISO 11133:2014/A1:2018; A2:2020.

Aerobiosis. Incubation at 20-25 °C, reading after 24-72 hours for bacteria and 3-5 days for yeasts and moulds

Microorganism

Growth

Candida albicans ATCC® 10231, WDCM 00054

Good

Escherichia coli ATCC® 8739, WDCM 00012

Good

Aspergillus brasiliensis ATCC® 16404, WDCM 00053

Good

Sterility control

Incubation 48 h at 30-35 °C and 48 h at 20-25 °C: NO GROWTH.

Check at 7 days after incubation in same conditions.

Bibliography

- AJELLO, L. (1957) Cultural Methods for Human Pathogenic Fungi J. Chron. Dis. 5:545-551.
- COLIPA (1997) Guidelines on Microbial Quality Management (MQM). Brussels.
- EUROPEAN PHARMACOPOEIA 7.0 (2011) 7th ed. § 2.6.13. Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. EDQM. Council of Europe. Strasbourg.
- GEORGE, L.K., AJELLO, L. & PAPAGEORGE, C. (1954) Use of Cycloheximide in the Selective Isolation of Fungi Pathogenic to Man. J. Lab. Clin. Med, 44 (422-428).
- HANTSCHKE, D. (1968) Mykosen, 11, (769-778).
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 13681 Standard. (1995). Enumeration of Yeasts and Moulds. Colony Count Technique.
- PAGANO, J. LEVIN, J.D. and TREJO, W. (1957-58) Diagnostic Medium for Differentiation of Species of Candida. Antibiotics Annual, 137-143.
- SABOURAUD, R. (1910) Les Teignes. Masson, Paris.
- USP 33 - NF 28 (2011) <62> Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. USP Corp. Inc. Rockville. MD. USA.