

Product :

BOLTON ENRICHMENT BROTH BASE



Revision date: 10/01/2019

Specification

Liquid culture medium used for the enrichment of *Campylobacter* from food samples according to the ISO standard.

Formula ^ in g/L	
Meat peptone	10.00
Lactalbumin hidrolysate	5.00
Yeast extract	
Sodium chloride	5.00
Sodium pyruvate	
Sodium metabisulfite	0.50
Sodium carbonate	0.60
a-Ketoglutaric acid	1.00
Haemin	0.01

Final pH 7.4 ±0.2 at 25 °C

Directions

Dissolve 13,8 g of the powder in 500 mL of distilled water, heating if necessary. Sterilize in the autoclave at 121° C for 15 minutes. Cool to $47-50^{\circ}$ C, add 25 mL of lysed horse blood aseptically, and the content of one vial of *Campylobacter* Bolton Selective Supplement (Ref. DSHB3033). Mix throughly. Dispense the complete medium into suitable containers. Note: If the enrichment broth has been prepared in advance, it should be kept for no more than 4 hours at ambient temperature or in the dark at $3 \pm 2^{\circ}$ C for not more than 7 days.

Description

Bolton Broth Base is intended for the enrichment of *Campylobacter* from food samples. Food processing and preservation injure *Campylobacter* cells and resuscitation steps by a double incubation in Bolton Broth encourages them to multiply and grow.

The meat peptone and lactalbumin hydrolysate supply the carbon and nitrogen for growth. Sodium chloride provides osmotic balance and the sodium carbonate neutralizes the acidity generated by the microbial growth. Yeast extract and ketoglutaric acid act as growth factors. Inclusion of sodium metabisulfite, sodium pyruvate and haemin neutralises toxic compounds that may form in the culture medium due to the action of oxygen action and avoid the need for a microaerobic atmosphere. Lysed blood is necessary to neutralize trimethoprim antagonists present in the medium.

The selectivity of the enrichment step is optimized with the Selective Supplement (Ref. DSHB3033): Vancomycin is active against Gram positive microorganisms. Cephoperazone is predominantly active against Gram negative bacteria. Trimethoprim acts against a wide variety of Gram positive and Gram negative cells and cycloheximide or amphotericin B are efficient fungicides.

Necessary supplements

Campylobacter Bolton Selective Supplement (Ref. DSHB3033)

Vial Contents:

Necessary amount for 500 mL of complete medium.

Distilled water (Solvent)

Technique

Introduce a quantity (mass or volume) into nine times its volume of Bolton Selective Enrichment Broth so as to obtain a test sample/medium ratio of 1:10 (w/v or v/v) and homogenize.

Bolton Selective Enrichment Broth does not require incubation in a microaerobic environment, but must be used in screw topped containers which are filled leaving a headspace of less than 20 mm, and have tightly closing caps.

Incubate the initial suspension at 37°C for 4-6 hours, then at 41,5°C for 44 ± 4 hours.

For the isolation and identification techniques, please, refer to ISO or BAM (Bacteriological Analytical Manual) methods.

^{*} Adjusted and /or supplemented as required to meet performance criteria

Reference: DSHB3033

Product :

BOLTON ENRICHMENT BROTH BASE



Technical data sheet

Quality control

Incubation temperature: 37°C ±1 / 41,5°C ±1 Incubation time: 5 h± 1 / 44 ±4 h

Inoculum: Practical range 100±20 CFU. Min. 50 CFU (Productivity) / 104-106 CFU (Selectivity) according to ISO

11133:2014. Microaerobic atmosphera

Microorganism Growth Remarks

Campylobacter jejuni ATCC® 29428 Good recovery in CCD ≥ 10 CFU Grey-flat-humid colonies, sometimes metallic shine

Escherichia coli ATCC® 8739 Inhibited in TSA -

Escherichia coli ATCC® 8739 Inhibited in TSA Proteus mirabilis ATCC® 29906 Inhibited in TSA -

ATCC[®] 29428 + 8739 +29906 Good recovery in CCD ≥ 10 CFU Grey-flat-humid colonies, sometimes metallic shine

References

- BAYLIS, C.L., (editor) (2007) Manual of Microbiological Methods for the Food and Drinks Industry. 5th ed. Method 3.3.1:2007. CCFRA. Chipping Campden. UK.
- · BOLTON, F.J. (2000) Methods for isolation of campylobacters from humans, animals, food and water. In "The increasing incidence of human campylobacteriosis" Report and Proceedings of a WHO Consultation of Experts. Copenhagen Denmark 21-25 November 2000, WHO/CDS/ CSRAPH 2001. p. 87-93.
- BOLTON, F.J., D. COATES, P.M. HINCHCLIFFE & L. ROBERTSON (1983) Comparison of selective media for isolation of Campylobacter jejuni/coli. J. Clin. Pathol. 36:78-83.
- · BOLTON, F.J., D. COATES & D.N. HUTCHINSON (1984) The ability of Campylobacter media supplements to neutralize photochemically induced toxicity and hydrogen peroxide. J. Appl. Bacteriol. 56:151-157.
- · CORRY, J.E.L., H. IBRAHIM ÁTABAY, S.J. FÓRSYTHE & L.P. MANSFIELD (2003) Culture Media for the isolation of campylobacters, helicobacters and arcobacters. In "Handbook of Culture Media for Food Microbiologists". J.E.L. Corry et al. (Eds.) Elsevier Science B.V. Amsterdam.
- · DOYLE, M.P. & D.J. ROMAN (1982) Recovery of Campylobacter jejuni and C. coli from inoculated foods by selective enrichment. Appl. Environm. Microbiol. 43:1343-1353.
- · FDA (Food and Drug Adminstrations) (1998) Bacteriological Analytical Manual. 8th ed. Revision A. AOAC International. Gaithersburg. Maryland. USA.
- · HUNT, J.M., C. ABEYTA & T. TRAN (1998) Campylobacter. In: FDA BAM 8th ed. (revision A) 7.01-7.027 AOAC International. Gaithersburg. MD. USA.
- · ISO 10272-1 Standard (2017) Microbiology of the food chain Horizontal Method for detection and enumeration of Campylobacter spp. Part 1: Detection method.
- · ISO 10272-2 Standard (2017) Microbiology of the food chain Horizontal Method for detection and enumeration of Campvlobacter spp. Part 2:Colony count-tecnique.
- . ISO 11133:2014. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- · STERN, N.J., J.E. LINE & H.C. CHEN (2001) Campylobacter in "Compendium of methods for the Microbiological Examination of Foods" 4th ed. F.P. Downes & K. Ito (Eds.) APHA, Washington. DC. USA.

Storage

For laboratory use only. Keep tightly closed, away from bright light, in a cool dry place (+4 °C to 30 °C).

Technical data sheet - page 2 of 2 Revision date: 10/01/2019